

Summary of Compliance - All Data

Compliance Statistics

Date Range	2/27/2005 - 3/4/2005 (6 days)
Days with Device Usage	6 days
Days without Device Usage	0 days
Percent Days with Device Usage	100 %
Cumulative Usage	48 hrs. 9 mins.
Maximum Usage (1 Day)	8 hrs. 58 mins.
Average Usage (All Days)	8 hrs. 2 mins.
Average Usage (Days Used)	8 hrs. 2 mins.
Minimum Usage (1 Day)	6 hrs. 46 mins.
Percent of Days with Usage >= 4 hours	100 %
Percent of Days with Usage < 4 hours	0 %

RemStar Auto Statistics

Auto-CPAP Mean Pressure	8.5 cm H2O
Auto-CPAP Peak Average Pressure	11.0 cm H2O
Average Device Pressure <= 90% of Time	10.0 cm H2O
Average Time in Apnea Per Day	2 secs.
Average Time in Large Leak Per Day	0 mins.